OneLight™ Rechargeable Light Source

Instructions for Use



℞ ONLY

bionix®

INTENDED USE
The intended use of the OneLight™ Rechargeable Light Source and the Original Light Source is to provide illumination during examination, inspection and treatments of patient(s).

INDICATIONS
The OneLight™ Rechargeable Light Source and Original Light Source are indicated for use on all patient populations for any medical procedures requiring illumination.

CONTRAINDICATIONS
There are no absolute contraindications for use of the OneLight™ Rechargeable Light Source.

RESIDUAL RISKS
Risks associated with the use of this product have been reduced as far as possible, but the product cannot completely eliminate potential patient or user harm arising from the following:

- Harm from mechanical hazards
- Harm from misuse or use errors
- Harm from unanticipated origins

⚠ WARNING
Cross-contamination risk. Do not reuse disposable instruments, e.g. curettes, as this may spread contamination from one patient to another patient. Discontinue curettage immediately if bleeding, irritation or other trauma to the ear canal or tympanic membrane occurs.

CHARGING

To charge the OneLight™ Rechargeable Light Source:

1. Connect the provided charging cable to the USB-C port on the device and plug the other end into the provided charging block.

   ⚠CAUTION: Use only the charging cable and block provided with this product.
2. Plug the charging block into a power outlet.
3. While charging, the OneLight™ Rechargeable Light Source power button will glow red.
   Once fully charged, the red light will turn off.
4. Note: The OneLight™ Rechargeable Light Source will not function while charging. Ensure the device is fully charged and disconnected from the power source before use.

Battery Specifications	
Estimated Time to Fully Charged Battery	~105 minutes
Estimated Continuous Use Run Time – from a fully charged battery	~120 minutes

INSTRUCTIONS FOR USE

Initial Setup

Before first use, remove the OneLight™ Rechargeable Light Source from its packaging and **fully charge** it.

Assembly

1. Attach the OneLight™ Rechargeable Light Source to a Lighted Bionix Instrument (e.g., curette, lighted forceps, etc.).
2. Align the instrument's end with the threaded opening of the OneLight™ Rechargeable Light Source, ensuring the pegs are properly positioned.
3. **Turn clockwise** until secure. Do not overtighten.
4. Attach the Magnification Lens – Snap the Magnification Lens onto the neck of the OneLight™ Rechargeable Light Source; ensure the top of the lens angles toward the tip of the lighted instrument.

Operation

1. To Turn ON: Press the button in the center of the OneLight™ Rechargeable Light Source.
2. To Turn OFF: Press the button again.

Disassembly and Storage

1. Rotate the disposable tool counterclockwise to detach it from the OneLight™ Rechargeable Light Source.
2. Discard disposable curette after use.
3. Return the OneLight™ Rechargeable Light Source to the charger as needed or store in a clean, dry environment.

LightLock Assembly & Disassembly

Align tongue depressor or the receiving tab on the OtoClear Wand with LightLock slot on opposite side of button. Slide tongue depressor or OtoClear Wand into the LightLock slot. To Disassemble, pull or push the tongue depressor or OtoClear Wand out of the LightLock slot.

CLEANING
Detailed cleaning and disinfection instructions can be found on the Bionix website under Instructions on the product page.

⚠ **WARNING:** Do **not** immerse the OneLight™ Rechargeable Light Source in any liquid.

DISPOSAL PROCEDURE
This device contains a Lithium-Ion battery and must be disposed of properly in accordance with local, state and federal regulations. Do not incinerate, crush, puncture or expose it to high temperatures.

To safely dispose of the battery:
- Follow your local electronic waste (e-waste) disposal guidelines.
- Take the device to an authorized battery recycling facility or an approved electronics recycler.
- Do not dispose of it in regular trash, municipal waste or a landfill.

For more information on proper disposal, contact your local waste management authority or visit **www.epa.gov/recycle/used-lithium-ion-batteries**

	Manufacturer		Do Not Use if Package is Damaged
	Date of manufacture		Do Not Reuse Single-Use Device
	Lot Code		Prescription Only or "For Use by or on the order of a licensed medical professional"
	Reorder Number		Consult Instructions For Use (IFU)
	Caution		Warning

Designed and Assembled in the USA



©2026 Bionix LLC | 1670 Indian Wood Circle
Maumee, Ohio 43537 | United States
Phone: 800.551.7096 (+1 419.727.8421)
Fax: 800.455.5678 (+1 419.727.8426)
Email: bionix@bionix.com (international@bionix.com)
www.bionix.com | RM95-0102 | Rev. 03